



Etac AS, Etac Supply Gjøvik  
Hadelandsveien 2  
2816 Gjøvik  
NORWAY  
Telephone: +47 40001004

## EC DECLARATION OF CONFORMITY

Hereby declare that the following product:

Molift Mover 180

and that the accessoires used only together with this product,  
are CE marked and complies with the Council Directive concerning medical devices 93/42/EEC,  
and according to this classified as medical equipment Class I.  
Annex IX rules that all non invasive devices are in Class I.

Expiration date: 15<sup>th</sup> June 2022

Product classification:

- 12 ASSISTIVE PRODUCTS FOR PERSONAL MOBILITY
- 12.36 ASSISTIVE PRODUCTS FOR LIFTING PERSONS
- 12.36.03 MOBILE HOISTS FOR TRANSFERRING A PERSON IN SITTING POSITION WITH SLING SEATS

Usage: Reusable medical device  
Sterility: Non  
Classification: MDD Class 1

Molift Mover 180 is manufactured in conformity with the following national or international standards that also might implement a harmonised standard:

EN 10535:2006 Hoists for the transfer of disabled persons - Requirements and test methods  
EN 14971:2012 Medical equipment - Risk analysis  
EN 60601-1 Medical electrical equipment  
EN 60601-1-2 Electromagnetic compatibility  
EN 60601-11 Medical electrical equipment used in the home healthcare environment  
RoHS 2 - Directive 2011/65/EU

Title: Quality Manager  
Name: Åsmund Amlien  
Company: Etac AS, Etac Supply Gjøvik

08.08.2017  
Date

Signature

---